



## ASSAM INDUSTRIAL DEVELOPMENT CORPORATION LIMITED

(A Government of Assam Undertaking)

R.G.Baruah Road, Guwahati-781 024

Ph. : 0361-2201215, 2202216, Fax : 0361-2200060

CIN: U75112AS1965SGC001246

Queries reply on supply, installation, commissioning, and training of food quality control lab equipment for RFTL at IID-Centre, Nalbari, Assam

The following replies have been recommended by the Internal Technical Committee meeting of AIDC held on 30.07.2018 in favour of the questionnaire.

Sl. No.	Page / Reference.	Reference of query	Query by Bidder	Reply by AIDC
1.	Instruction to Bidders. Clause No.1.11	<b>A.Dr. Anuranjan Pandeya</b> , Regional Manager, <b>Scientific and Digital Systems, IDA House, Sector 4, R.K. Puram, New Delho-110022</b> Moblie No. +91-11-26170781,26165355  <b>Email vide dated 24.07.2018</b>	The party stated most of instruments are manufactured by different manufactures located in difference part of world which are imported by their exclusive representatives in India. Finding a representative (supplier) who can offer all the instruments from the tender is nearly impossible. Even if anyone Company agrees to supply all these equipments, they will have no technical expertise and hence the usage and the quality will be	Clause No. 1.11 of Section-II "Instruction to Bidders" having in RFP Document will be applicable.

			<p>largely compromised.  <b>Hence the party requested us to change the terms and conditions of the tender and allow the suppliers to bid for single or few or as many equipments they want. Also, the earnest money may be fixed for individual equipments rather for the complete tender. This will allow many suppliers to participate.</b></p>	
2	<p><b>Sec-IX  Name of Equipment:  GC-MS-MS</b></p>	<p><b>Rounak Ganguli,</b>  Management Trainee,  Chromatography &amp; Mass Spectrometry Division, Thermo Fisher Scientific India Pvt. Ltd.  Ph.. no: 8910361016</p> <p><b>Email vide dated 25.07.2018</b></p>	<p>The party has proposed a new modernized specification for the lab equipment of the <b>Triple Quadrupole GC-MS/MS</b> system instead of <b>Iron Trap GC-MS</b> system mentioned in the RFP for generalization of the specification.</p>	<p>The Corporation can agree to the proposal if MoFPI has already approved this technology of <b>Triple Quadrupole GC-MS/MS</b> system substituting <b>Iron Trap GC-MS</b> system.</p>